

## Applicant Initiated Interview Request Form

Application No.: 10/811,733 First Named Applicant: Dr. Jeffrey A. Klein  
 Examiner: Bertheaud, Peter Art Unit: 3746 Status of Application: Pending

**Tentative Participants:**

- |                                 |                               |
|---------------------------------|-------------------------------|
| (1) <u>Dr. Jeffrey A. Klein</u> | (2) <u>James C. Yang</u>      |
| (3) <u>Kit M. Stetina</u>       | (4) <u>Examiner Bertheaud</u> |

**RECEIVED**  
**CENTRAL FAX CENTER**  
**APR 30 2008**

Proposed Date of Interview: TBD Proposed Time: TBD AM/PM

**Type of Interview Requested:**

- (1)  Telephonic      (2)  Personal      (3)  Video Conference

Exhibit To Be Shown or Demonstrated:  YES  NO

If yes, provide brief description: \_\_\_\_\_

### Issues To Be Discussed

Issues (Rej., Obj., etc)	Claims/ Fig. #s	Prior Art	Discussed	Agreed	Not Agreed
(1) <u>Rej.</u>	<u>1 (Now 29)</u>	<u>Holtorf</u>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
(2) <u>Rej.</u>	<u>4 (Now 32)</u>	<u>Reimels</u>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
(3) <u>Rej.</u>	<u>12 (Now 40)</u>	<u>Burbank</u>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
(4) <u>Rej.</u>	<u>15 (Now 43)</u>	<u>Burbank</u>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

Continuation Sheet Attached

**Brief Description of Argument to be Presented:**

See attached.

An interview was conducted on the above-identified application on \_\_\_\_\_.  
**NOTE: This form should be completed by applicant and submitted to the examiner in advance of the interview (see MPEP § 713.01).**  
**This application will not be delayed from issue because of applicant's failure to submit a written record of this interview. Therefore, applicant is advised to file a statement of the substance of this interview (37 CFR 1.133(b)) as soon as possible.**

James C. Yang  
 Applicant/Applicant's Representative Signature  
James C. Yang  
 Typed/Printed Name of Applicant or Representative  
54,556  
 Registration Number, if applicable

\_\_\_\_\_  
 Examiner/SPE Signature

This collection of information is required by 37 CFR 1.133. The information is required to obtain or retain a benefit by the public which is to file (and by the USPTO to process) an application. Confidentiality is governed by 35 U.S.C. 122 and 37 CFR 1.11 and 1.14. This collection is estimated to take 21 minutes to complete, including gathering, preparing, and submitting the completed application form to the USPTO. Time will vary depending upon the individual case. Any comments on the amount of time you require to complete this form and/or suggestions for reducing this burden, should be sent to the Chief Information Officer, U.S. Patent and Trademark Office, U.S. Department of Commerce, P.O. Box 1450, Alexandria, VA 22313-1450. DO NOT SEND FEES OR COMPLETED FORMS TO THIS ADDRESS. SEND TO: Commissioner for Patents, P.O. Box 1450, Alexandria, VA 22313-1450.

RECEIVED  
CENTRAL FILE CENTER

APR 30 2008

**ISSUES TO BE DISCUSSED**

#1 – Claims 29: Holtorf does not appear to teach a foot pedal delivering one pulse of air to a control air inlet in the console. Rather, Holtorf appears to suggest that a signal is generated by the foot pedal related to the magnitude of foot pedal depression. See Col. 4, Lines 5-6 of Holtorf). Moreover, in Claim 1 (now Claim 29), the claim is not limited to one pulse of air.

#2 – Claims 32: Reimels does not appear to teach adjusting a fluid flow rate based on a duration of foot pedal depression. Rather, it appears that Reimels adjusts the fluid flow rate based on the amount of depression. (See Col. 5, Line 51 – Col. 6, Line 9 of Reimels).

#3 – Claim 40: Burbank does not appear to disclose a sensor “in contact with” a flexible tubing. The Burbank reference shows an ultrasonic sensor which depends on sound waves and not on mechanical contact between the sensor and the flexible tubing. Also, Figure 4 of Burbank shows a gap between the ultrasonic sensor 30 and the flexible tubing.

#4 – Claim 43: Burbank does not show a force sensor “in contact with” rollers. Support for a force sensor is found in paragraph 10 of the instant application.

T:\Client Documents\KLEIN\084BC\Issues to be Discussed